



中国认可
国际互认
校准
CALIBRATION
CNAS L9002



22BA000280001



安正计量检测有限公司

ANZHENG METROLOGY AND TEST CO., LTD.

校准证书

Calibration Certificate

证书编号:

BA20223460091

Certificate No.

委托单位

万霖消防技术有限公司

Client

委托地址

北京市通州区西集镇国防路43号A座595

Address

器具名称

无线远程数显液位计

Description

制造厂家

Manufactory

型号 / 规格

WANLIN-907

Type or size

器具编号

YWJ-01

Number of sample

校准结果

所校准项目符合技术要求

Calibration Conclusion



批准人:

Authorized by

李明祥

核验员:

Checked by

雷飞

校准员:

Calibrated by

汪俊峰

接收日期

Date of Receipt

2022

年

01

月

19

日

校准日期

Date of Calibration

2022

年

01

月

19

日

发布日期

Date of Publication

2022

年

01

月

19

日

地

址:

福建省福州市仓山区仓山科技园1区02号1#楼101室

ADD: Room 101, Building 1#, No 02, District 1, cangshan Science and Technology Park, cangshan district, Fuzhou city, Fujian province

邮编(Post Code):

350026

服务电话(Tel):

0591-88030652

传真(Fax): 0591-83591276

网址(Web): www.fjazjl.com



校准结果

Result of Calibration

1、校准用介质：

Calibration media:

2、外观：

Appearance:

3、量程：

Scope of flow:

4、流量示值误差校准：

Calibration of flow indicator error:

被校仪表示值	标准器示值	示值误差	允许误差	不确定度	结论
Instrument value	Standard Value	Indication Error	MPE	Uncertainty	Conclusion
mm	mm	mm	mm	$U_{rel}(k=2)$	(P/F)
100	100.00	0.00	± 3.0	0.9mm	P
1000	1001.27	-1.27			P
2000	2001.67	-1.67			P
3000	3001.88	-1.88			P
4000	4002.18	-2.18			P
5000	5002.24	-2.24			P

备注：

Notes:

1、为确保校准结果的可信度，通常情况下建议再校周期为12个月。

To ensure the credibility of the calibration results, it is generally recommended to have a recalibration period of 12 months.

(以下空白)

(The below is blank)